Anigen Rapid

CIRD-3 Ag

General Information

Purpose : Detection of Canine Infectious Respiratory Disease (CDV, CAV and CIV)

Principle : Immunochromatographic assay

Specimen : CDV Ag - Conjunctiva, CAV & CIV Ag - Nasal fluid

Materials provided :

- Anigen Rapid CIRD-3 test device
- Sample collection containing assay diluents
- Disposable swab
- Disposable dropper

Special Features :

- Simultaneous detection of CDV, CAV and CIV
- with one specimen preparation
- One-step testing procedure : Quick & Accurate
- High sensitivity and specificity



Test Procedure



1 Collect a sample from conjunctiva and nasal cavity



2 Insert the swab into the sample collection tube and stir the swab at least 10 times






3 Discard the swab while squeezing the swab against the wall of the tube

4 Take the supernatant by using the disposable dropper provided



Add 4 drops into each the sample hole on the test device

Interpretation



Ordering Information

Cat. No.	Description	Туре	Pack size
RG11-09	CIRD-3 Ag	Dual Device	10 Tests/Kit