



Instructions for Use

380292-A

English

For Veterinary Research Use Only

VET-SAA-QC-Low

REF V-SZ91
REF G-SZ94

INTENDED USE

The product is intended for use as an internal quality control fluid in laboratories making use of automated analyzers to measure Serum Amyloid A (SAA) for animal.

REAGENTS

Reagents are stable until the date printed on the label assuming the container remains unopened at a storage temperature of 2-8°C.

WARNINGS AND PRECAUTIONS

- For veterinary research use only.
- Do not freeze product.
- Do not use expired product.
- Do not dissolve with other than the dedicated Lysis buffer.
- Air bubbles on the surface of the product after being poured into sample cups can result in error measurement, and should be removed.
- No human-derived ingredients are contained in this product, however, the risk of infection and the possible existence of other pathogens cannot be ruled out completely. Handle it as carefully as an animal sample.
- The product should be used with VET-SAA Calibrator Set (REF V-SZ90).
- Use the reagents as quickly as possible after reconstitution.
- Dispose of used reagents and containers as infectious waste in accordance with applicable regulations.
- The product contains Guanidinium chloride (CAS No.50-01-1) and ProClin 300 (CAS No.55965-84-9) which are classified by applicable 'The Classification, Labelling and Packaging (CLP) Regulation' as Skin and Eye Irritant and Acute toxicity by Oral. If the reagent contacts with eyes, mouth or skin, rinse it out with large volumes of running water, and perform other required first aid. If necessary, seek medical attention.



Warning

H302	Harmful if swallowed.
H315	Causes skin irritation.
H317	May cause an allergic skin reaction.
H319	Causes serious eye irritation.
P261	Avoid breathing dust/fume/gas/mist/vapours/spray.
P264	Wash hands thoroughly after handling.
P270	Do not eat, drink, or smoke when using this product.
P272	Contaminated work clothing should not be allowed out of the workplace.
P280	Wear protective gloves/protective clothing/eye protection/ face protection.
P301+ P312	IF SWALLOWED: Call a POISON CENTER or doctor/ physician if you feel unwell.
P330	Rinse mouth.

P302+ P352	IF ON SKIN: Wash with plenty of soap and water.
P305+ P351+ P338	IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.
P321	Specific treatment.
P332+ P313	If skin irritation occurs: Get medical advice/attention.
P333+ P313	If skin irritation or rash occurs: Get medical advice/attention.
P337+ P313	If eye irritation persists: Get medical advice/attention.
P362+ P364	Take off contaminated clothing and wash it before reuse.
P501	Dispose of contents/container in accordance with local disposal regulations.

PROCEDURE

- Gently remove the rubber stopper from the vial. Add 1.0mL of Lysis buffer that brought back to a room temperature with a pipette.
- Seal the vial with the rubber stopper, and mix by gently inverting the vial several times. Reconstitute the control at least 15 minutes before use.
- Dispense an appropriate quantity of the reconstituted control into a sample cup and then operate the automatic analyzer by following to its instructions.

PRODUCT CODE, PRODUCT NAME & STORAGE

Product code	Product name	Contents	Storage
V-SZ91	VET-SAA-QC-Low	Control : 5 × 1mL Lysis buffer : 1 × 12mL	2-8°C

CONCENTRATION

Range of quality control is stated on the vial label.



EIKEN CHEMICAL CO., LTD.
4-19-9, Taito, Taito-ku, Tokyo 110-8408, Japan

(Date of issuance: 2019-10-01)